

VALTROL-P

Suero Control Patológico LOTE 220601 VENCE 11-2025

| Componente | Método | Valor | Intervalo | 1S | Unidad |
|------------------------------|--------------------------------------|--------|-----------------|--------|--------|
| ACIDO URICO | UOD - PAP DCBS | 9,40 | 7,43 - 11,37 | 0,658 | mg/dl |
| ALBUMINA | BCG | 3,24 | 2,56 - 3,92 | 0,227 | g/dl |
| ALT/ALAT/GPT | IFCC sin Piridoxal Fosfato (37°C) | 142,70 | 112,73 - 172,66 | 9,989 | U/l |
| AST/ASAT/GOT | IFCC sin Piridoxal Fosfato (37°C) | 151,46 | 119,65 - 183,26 | 10,602 | U/l |
| AMILASA | CNPG3 - líquida | 312,29 | 246,71 - 377,87 | 21,860 | U/l |
| BILIRRUBINA DIRECTA | Diazo | 1,822 | 1,44 - 2,2 | 0,128 | mg/dl |
| | Walters-Gerarde/Malloy-Evelyn (37°C) | 1,714 | 1,2 - 2,23 | 0,171 | mg/dl |
| BILIRRUBINA TOTAL | Diazo | 5,450 | 4,31 - 6,59 | 0,381 | mg/dl |
| | Walters-Gerarde/Malloy-Evelyn (37°C) | 5,151 | 3,61 - 6,7 | 0,515 | mg/dl |
| CALCIO | Arsenazo | 13,63 | 10,77 - 16,49 | 0,954 | mg/dl |
| | | 6,82 | 5,38 - 8,25 | 0,477 | mEq/l |
| | | 3,41 | 2,69 - 4,12 | 0,239 | mmol/l |
| | o-cresolftaleína-complexona | 14,02 | 11,08 - 16,97 | 0,982 | mg/dl |
| | | 7,01 | 5,54 - 8,48 | 0,491 | mEq/l |
| | | 3,51 | 2,77 - 4,24 | 0,245 | mmol/l |
| COLESTEROL TOTAL | CHOD-PAP | 300,77 | 237,6 - 363,93 | 21,054 | mg/dl |
| COLESTEROL HDL (*) | Homogéneo – Directo | 92,63 | 73,18 - 112,08 | 6,484 | mg/dl |
| CREATINQUINASA (CK) | IFCC (37°C) | 568,60 | 449,19 - 688,01 | 39,802 | U/l |
| CREATININA | Enzimática | 4,72 | 3,73 - 5,71 | 0,330 | mg/dl |
| FOSFATASA ALCALINA | IFCC – LS (líquido) (37°C) | 346,52 | 273,75 - 419,29 | 24,257 | U/l |
| | DGKC (37°C) | 446,70 | 352,9 - 540,51 | 31,269 | U/l |
| FOSFORO INORGANICO | Molibdato UV sin blanco muestra | 6,47 | 5,11 - 7,82 | 0,453 | mg/dl |
| γ-GLUTAMILTRANSFERASA (GGT) | Szasz (37°C) | 186,19 | 147,09 - 225,29 | 13,033 | U/l |
| GLUCOSA | GOD-PAP | 291,88 | 230,59 - 353,18 | 20,432 | mg/dl |
| LACTATO DESHIDROGENASA (LDH) | IFCC - LS (37°C) | 389,55 | 307,75 - 471,36 | 27,269 | U/l |
| PROTEINA TOTAL | Biuret | 5,19 | 4,1 - 6,27 | 0,363 | g/dl |
| TRIGLICERIDOS | GPO-PAP | 253,23 | 200,05 - 306,4 | 17,726 | mg/dl |
| UREA | Ureasa UV cinética | 121,58 | 96,05 - 147,11 | 8,510 | mg/dl |

(*) El valor obtenido para Colesterol HDL considera la utilización del Calibrador Colesterol HDL VALTEK código # 070-120